

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 13 July 2016
to 27 October 2018
Issued on 25 July 2016



Schedule of Accreditation

Accreditation Scheme for Testing/Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 007-01

Lanka Calibration Services (Pvt) Ltd
No. 27/14A
Rosmead Place
Colombo 07.

Scope of Accreditation: Performing thermal calibrations as per the calibration methods given in the schedule of accreditation.

The laboratory is accredited for the calibration as given page 02.

Sl No	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	Calibration Measurement Capability (Approximately at 95% Confidence Level)	Location
01	Verification of temperature enclosures Parameter: temperature	LCS/TM/12/Rev.0 (Based on Temperature mapping of storage areas Technical supplement to WHO Technical Report Series, No. 961, 2011 Annex 9: Model guidance for the storage and transport of time and temperature-sensitive pharmaceutical products : January 2014 by World Health Organization 2014)	-35 °C 0 °C +80 °C	-	1.1 °C 0.7 °C 1.2 °C	Site
02	Calibration of IR Thermometer Parameter: temperature	LCS/TM/10/Rev .0 (Based on ASTM E2847-14)	Ambient (20 °C) +140 °C	-	0.5 °C 0.8 °C	In-House

Deputy Director (Accreditation)
Sri Lanka Accreditation Board for Conformity Assessment