

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 30 October 2019
To 30 November 2020
Issued on 30 October 2019



Schedule of Accreditation

Accreditation Scheme for Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 015-01

**Biomedical Engineering Laboratory,
Wickramarachchi Institute of Speech and Hearing,
J M Wickramarachchi & Company Private Limited
No. 436, Delkanda,
Nugegoda.**

Scope of Accreditation: : Performing calibration of pure tone audio meters as per acoustic method appearing in this Schedule.

The laboratory is accredited for the following tests as per given in the page 02 of 02.

SI	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	Calibration Measurement Capability (Approximately at 95% Confidence Level)	Location
Sound and Vibrations						
01	Sound and Vibrations	ISO 389-1 Calibration of sound pressure level for pure tones and supra aural headphones IEC 60645-1 Equipment for pure tone and speech audiometry	-5 – 120dB	5dB	±3	Laboratory
		ISO 389-3 Calibration of vibratory force level for pure tones and bone vibration IEC 60645-1 Equipment for pure tone and speech audiometry	-5 – 70dB			

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment