



Valid from 25 September 2017  
to 24 September 2020  
Issued on 25 September 2017

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



ISO/IEC 17025  
TL 029-02

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment  
Accreditation Number: TL 029-02

### **Physical Testing Laboratory DPL Premier Gloves Limited Block B, BEPZ Walgama, Malwana**

**Scope of Accreditation:** Performing Mechanical testing on Rubber Products as per the test methods appearing in this schedule

The laboratory is accredited for the following tests.

Sl No	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	<b>Rubber Products</b> (Protective gloves against mechanical risks)	Abrasion Resistance	BS EN388: 2016 section 6.1	1-8800 no of rubs
		Blade Cut Resistance	BS EN388: 2016 section 6.2	1.2 -21.5 Index
		Blade Cut Resistance	BS EN388: 2016 section 6.3 BS EN ISO 13997: 1999	2.2 - 33.0 N
		Tear Resistance	BS EN388: 2016 section 6.4	0 - 81.0 N
		Puncture Resistance	BS EN388: 2016 section 6.5	0 - 161.0N
		Tensile Strength Measurement	ASTM D 412 : 2016	1 - 40MPa

Deputy Director (Accreditation)  
Sri Lanka Accreditation Board for Conformity Assessment