



As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.

Valid from 16 June 2021  
to 16 May 2024  
Issued on 17 May 2021



ISO/ IEC 17025  
CL 012-01

## Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment  
Accreditation Number: CL 012-01

### WAGA Calibration Services (Pvt) Ltd

275B, Railway Road  
Maharagama

**Scope of Accreditation:** Performing Mechanical calibration (Force, Pressure, Mass, Dimension, Volume), Electrical, Thermal, Optical and Time & Frequency and Machine performance verification as per the calibration methods appearing in the schedule.

The laboratory is accredited for the following calibrations.

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
<b>1. Electrical</b>						
1.1	Analog and Digital Electrical measuring equipment / Indicators	DC Voltage	WAGA/CM/30	2 mV to 200 mV	0.06 mV	Laboratory
				0.2 V to 20 V	0.06 mV	
		AC Voltage (rms)		20 V to 200 V	0.2 mV	Laboratory
				200 V to 1050 V	6.0 mV	
		DC Current		-10 mV to 12 V	0.06 mV	Site
				1 mV to 200 mV (20 Hz to 20 kHz)	0.6 mV	Laboratory
		AC Current (rms)		0.2 V to 20 V (20 Hz to 20 kHz)	0.6 mV	
				Digital Frequency	20 V to 200 V (20 Hz to 20 kHz)	1.3 mV
		200 V to 1000 V (40 Hz to 1 kHz)			0.059 V	
		Digital Frequency		20 μA to 200 mA	0.6 mA	Laboratory
200 mA to 22 A	0.6 mA		Site			
Digital Frequency	1 mA to 22 mA	0.6 mA	Site			
	10 μA to 200 mA (20 Hz to 5 kHz)	0.6 mA	Laboratory			
Digital Frequency	200 mA to 20 A (20 Hz to 500 Hz)	0.6 mA	Laboratory			
	0.1 Hz to 10 MHz	0.058 Hz	Laboratory			
Digital Frequency	100 Hz to 50 kHz	0.082 Hz	Site			

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
1.1	Analog and Digital Electrical measuring equipment / Indicators	Inductance	WAGA/CM/30	1 mH to 1.0 H (at 1 kHz)	0.01 mH	Laboratory
		Resistance		1 $\Omega$ to 10 M $\Omega$	0.6 m $\Omega$	Laboratory
				100 M $\Omega$	0.01 M $\Omega$	
				1 G $\Omega$	0.03 G $\Omega$	
		1 $\Omega$ to 4.0 k $\Omega$		0.011 $\Omega$	Site	
Capacitance	1 nF to 1 $\mu$ F (1 kHz) 1 $\mu$ F to 50 $\mu$ F (1 kHz)	0.006 nF 0.006 $\mu$ F	Laboratory			
Power Factor	-1 to 1 (Lagging & Leading)	0.006 %	Laboratory			
1.2	Analog and Digital Electrical Equipment with sources	DC Voltage	WAGA/CM/030	0 mV to 100 mV	0.006 mV	Laboratory
				100 mV to 1 V	0.023 mV	
				1 V to 10 V	0.058 mV	
				10 V to 100 V	0.47 mV	
				100 V to 900 V	12 mV	
		AC Voltage (rms)		1 mV to 30 V	0.32 mV	Site
				30 V to 60 V	0.014 V	
				60 V to 600 V	0.13 V	
		DC Current		20 mV to 100 V	0.35 mV	Laboratory
				100 V to 900 V (3 Hz to 300 kHz)	0.69 V	
				60 mV to 60 V	0.36 mV	Site
		AC Current (rms)		60 V to 600 V (45 Hz to 1 kHz)	0.69 V	
				0 mA to 3 A	1.2 $\mu$ A	Laboratory
				3 A to 9 A	2.4 mA	
		Digital Frequency		0 A to 9 A	2.5 mA	Site
0.1 A to 3 A (3 Hz to 5 kHz)	0.24 mA		Laboratory			
3 A to 9 A (45 Hz to 500 Hz)	4.7 mA					
Resistance	0.6 A to 9 A	4.7 mA	Site			
	100 Hz to 100 kHz	0.0015 Hz	Laboratory			
	10 Hz to 50 kHz	0.082 Hz	Site			
Capacitance	0 $\Omega$ to 100 k $\Omega$	0.12 $\Omega$	Laboratory			
	100 k $\Omega$ to 10 M $\Omega$	4.7 $\Omega$				
	1 $\Omega$ to 4 k $\Omega$	1.5 m $\Omega$	Site			
1 $\mu$ F to 100 $\mu$ F	5.8 nF	Laboratory / Site				
1.3	Clamp Meters	DC Current	WAGA/CM/027	20 $\mu$ A to 20 A	1.2 mA	Laboratory
				20 A to 100 A	0.12 A	
				100 A to 500 A	0.58 A	
				500 A to 1100 A	2.9 A	
AC Current	10 $\mu$ A to 20 A	1.2 mA	Laboratory			
	20 A to 100 A	0.12 A				
	100 A to 500 A	0.58 A				
	500 A to 1000 A	2.9 A				
AC Voltage	1 mV to 200 mV (20 Hz to 20 kHz)	0.6 mV	Laboratory			
	0.2 V to 20 V (20 Hz to 20 kHz)	0.6 mV				
	20 V to 200 V (20 Hz to 20 kHz)	1.3 mV				
	200 V to 1000 V (40 Hz to 1 kHz)	0.059 V				
DC Voltage	2 mV to 200 mV	0.06 mV	Laboratory			
	0.2 V to 20 V	0.06 mV				
	20 V to 200 V	0.2 mV				
	200 V to 1050 V	6.0 mV				

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
1.3	Clamp Meters	Resistance	WAGA/CM/027	1 Ω to 10 Ω 100 MΩ 1 GΩ	0.6 mΩ 0.01 MΩ 0.03 GΩ	Laboratory
1.4	Insulation Testers	Generation of insulation resistance	WAGA/CM/028	100 kΩ to 5 MΩ 10 MΩ to 100 MΩ 1 GΩ to 100 GΩ	0.02 kΩ 0.009 MΩ 5.9 MΩ	Laboratory
		Open Circuit Voltage (Measurement of Voltage)		0.5 kV to 2 kV 2 kV to 10 kV	5.9 V 5.8 V	
		Short Circuit Current (measurement of current)		0.5 mA to 2 mA 2 mA to 20 mA	0.006 mA 0,059 mA	
1.5	Multimeters	DC Voltage	WAGA/CM/017	2 mV to 200 mV 0.2 V to 20 V 20 V to 200 V 200 V to 1050 V	0.003 mV 0.002 mV 0.1 mV 1.0 mV	Laboratory
		AC Voltage		1 mV to 200 mV (20 Hz to 20 kHz) 0.2 V to 20 V (20 Hz to 20 kHz) 20 V to 200 V (20 Hz to 1 kHz) 200 V to 1000 V (40 Hz to 1 kHz)	0.001 mV 0.1 mV 1.2 mV 0.012 V	
		DC Current		20 μA to 200 mA 200 mA to 22 A	0.001 mA 0.024 mA	
		AC Current		10 μA to 200 mA (20 Hz to 5 kHz) 200 mA to 20A (20 Hz to 500 Hz)	0.06 μA 0.05 mA	
		Resistance		1 Ω to 10 MΩ 100 MΩ 1 GΩ	0.1 mΩ 0.01 MΩ 0.03 GΩ	
		Capacitance		1 nF to 1 μF (1 kHz) 1μF to 50 μF (1 kHz)	0.006 nF 0.006 μF	
		Digital frequency		0.1 Hz to 100 MHz	0.058 Hz	
1.6	Calibration of pH Meter	Simulation of Voltage (Indicator)	WAGA/CM/024	0 pH, 4 pH, 7 pH, 10 pH, 14 pH (-413 mV to +413 mV)	0.0006 pH	Laboratory / Site
		Standard Reference Buffer Solutions (Electrode)		pH 0 to 14	0.007 pH	
<b>2. Thermal</b>						
2.1	Dial Thermometer	Direct Comparison	WAGA/CM/029	-30 °C to 100 °C 100 °C to 200 °C 200 °C to 400 °C 400 °C to 600 °C	0.14 °C 0.21 °C 0.34 °C 0.90 °C	Laboratory
				50 °C to 600 °C	0.34 °C	

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location	
2.1	Temperature Indicators and Simulators by Electrical Simulation and Measurement	<b>Electrical simulation of Temperature indicators / controllers intended to be used with thermocouple</b>					
		Thermocouple Type					
		K	WAGA/CM/023	-200 °C to 1370 °C	0.68 °C	Laboratory / Site	
		J		-210 °C to 1200 °C	0.68 °C		
		T		-200 °C to 400 °C	0.68 °C		
		R		-50 °C to 1765 °C	0.69 °C		
		S		- 50 °C to 1760 °C	0.69 °C		
		B		300 °C to 1800 °C	0.70 °C		
		E		-200 °C to 1000 °C	0.68 °C		
		N		-200 °C to 1300 °C	0.68 °C		
		C		0 °C to 2315 °C	0.69 °C		
		L		-200 °C to 900 °C	0.68 °C		
		U		-200 °C to 600 °C	0.68 °C		
		<b>Electrical simulation of temperature indicators / controllers intended to be used with PRT / RTD</b>					
PT 100	WAGA/CM/023	-180 °C to 850 °C	0.36 °C	Laboratory / Site			
2.2	Temperature Indicators and Simulators by Electrical Simulation and Measurement	<b>Electrical simulation of Temperature simulator intended to be used with thermocouple</b>					
		Thermocouple Type					
		K	WAGA/CM/023	-200 °C to 1370 °C	0.17 °C	Laboratory / Site	
		J		-210 °C to 1200 °C	0.17 °C		
		T		-200 °C to 400 °C	0.17 °C		
		R		-50 °C to 1765 °C	0.48 °C		
		S		-50 °C to 1760 °C	0.48 °C		
		B		200 °C to 1800 °C	0.63 °C		
		E		-200 °C to 1000 °C	1.1 °C		
		N		-200 °C to 1300 °C	0.17 °C		
		C		0 °C to 2315 °C	0.21 °C		
		L		-200 °C to 900 °C	0.15 °C		
		U		-200 °C to 600 °C	0.16 °C		
		<b>Electrical simulation of temperature simulator intended to be used to be used with PRT/RTD</b>					
PT 100	WAGA/CM/023	-180 °C to 850 °C	0.58 °C	Laboratory / Site			
2.3	Temperature enclosures – Nine point	Multi-point temperature verification	WAGA/CM/020	30 °C to 300 °C	0.56 °C	Laboratory / Site	
2.4	Temperature enclosures – Three point	Three-point temperature verification	WAGA/CM/041	30 °C to 300 °C	0.68 °C	Laboratory / Site	

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
2.5	Muffle / Tube furnace	Temperature verification	WAGA/CM/045	100 °C to 400 °C 400 °C to 1100 °C	0.93 °C 1.5 °C	Site
2.6	Thermocouples	Direct comparison under controlled conditions	WAGA/CM/015	150 °C to 650 °C	1.5 °C	Laboratory
2.7	Liquid Baths	Performance verification in temperature	WAGA/CM/042	25 °C to 250 °C	0.31 °C	Laboratory / Site
2.8	Hygrometers	Direct comparison under controlled conditions (Temperature, Humidity)	WAGA/CM/047	10 to 40 °C	0.1 °C	Laboratory
				50 % to 90 % (@25 °C)	3.0 %	
<b>3. Pressure</b>						
3.1	Pressure transmitter / Transducer	Measurement of pneumatic pressure Measurement of hydraulic pressure	WAGA/CM/021	0 to 40 bar	0.047 mV/V/bar 0.0099 mA/A/bar	Laboratory
				0 to 700 bar	0.047 mV/V/bar 0.010 mA/A/bar	
				0 to 1000 bar	0.049 mV/V/bar 0.018 mA/A/bar	
<b>4. Mass</b>						
4.1	Weighing Scales	Linearity check with conventional mass values	WAGA/CM/044	0 kg to 150 kg 150 kg to 300 kg	0.93 g 4.0 g	Laboratory / Site
4.2	Determination of conventional Mass Value: Class F1 and below Class F1	Based on double substitution method (ABBA) OIML R 111 and MSL Technical Guide 7	WAGA/CM/003	<b>Class F1 and below F1</b>		Laboratory
				1 g	0.063 mg	
				2 g	0.063 mg	
				5 g	0.063 mg	
				10 g	0.063 mg	
				20 g	0.063 mg	
				50 g	0.065 mg	
				100 g	0.070 mg	
				200 g	0.078 mg	
				500 g	0.63 mg	
				1000 g	0.66 mg	
				2000 g	1.2 mg	
				5000 g	2.2 mg	
				<b>Class F2 and below F2</b>		
10 kg	5.6 mg					
20 kg	0.2 g					
<b>5. Dimension</b>						
5.1	Dial Gauge	Linear measurement by direct comparison	WAGA/CM/035	0 mm to 25 mm 25 mm to 50 mm	0.9 µm 6.0 µm	Laboratory
5.2	Steel Ruler	Linear measurement by direct comparison	WAGA/CM/014	0 mm to 1000 mm	0.58 mm	Laboratory
5.3	Vernier Caliper	Linear measurement by direct comparison	WAGA/CM/036	0.5 mm to 150 mm 150 mm to 500 mm 500 mm to 1000 mm (External calibration)	0.9 µm 8.8 µm 13 µm	Laboratory
				0.5 mm to 200 mm (Depth calibration)	0.9 µm	Laboratory

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
5.4	External Micrometer	Linear measurement by direct comparison	WAGA/CM/013	0.5 mm to 50 mm 50 mm to 100 mm 100 mm to 500 mm	1.0 µm 1.3 µm 2.0 µm	Laboratory
5.5	Height Gauge	Linear measurement by direct comparison	WAGA/CM/037	0.5 mm to 200 mm 200 mm to 300 mm 300 mm to 500 mm	8.8 µm 9.5 µm 10.0 µm	Laboratory
5.6	Feeler Gauge	Linear measurement by direct comparison	WAGA/CM/043	0.10 mm to 1 mm	2.9 µm	Laboratory
<b>6. Volume</b>						
6.1	Laboratory glassware	Determination of volume by gravimetric method in: One mark pipette Graduate pipette Burette Volumetric Flask Graduated measuring cylinder Graduated beaker Conical flask	WAGA/CM/038	0 to 5 ml	0.004 ml	Laboratory
				0 to 50 ml	0.009 ml	
				0 to 100 ml	0.009 ml	
				0 to 200 ml	0.014 ml	
				0 to 500 ml	0.32 ml	
				0 to 1000 ml	0.64 ml	
				0 to 2000 ml	1.3 ml	
<b>7. Force</b>						
7.1	Torque Measuring Devices	Torque measurement by direct comparison	WAGA/CM/040	0 to 50 Nm	0.88 Nm	Laboratory
				0 to 220 Nm	0.88 Nm	
				0 to 2200 Nm	0.88 Nm	
<b>8. Optical</b>						
8.1	Illuminance Meters	Illuminance measurement	WAGA/CM/039	0 Lux to 10 Lux 10 Lux to 20 Lux 20 Lux to 200 Lux 200 Lux to 1000 Lux 1000 Lux to 2000 Lux 2000 Lux to 10000 Lux 10000 Lux to 20000 Lux	3.2 % 3.1 % 2.5 % 2.2 % 2.6 % 3.7 % 6.0 %	Laboratory
<b>9. Time and Frequency</b>						
9.1	Tachometers	Non-contact Type by Electrical Pulse	WAGA/CM/019	6 rpm to 1000 rpm 1000 rpm to 10000 rpm 10000 rpm to 100000 rpm	0.007 rpm 0.06 rpm 0.6 rpm	Laboratory
9.2	Rotating Machines (including centrifuges)	Speed	WAGA/CM/018	6 rpm to 1000 rpm	1.6 rpm	Laboratory / Site
				1000 rpm to 99,999 rpm	6.2 rpm	
<b>10. Performance Verification</b>						
10.1	Colour Fastener Testers	Temperature	WAGA/CM/032	30 °C to 100 °C	0.28 °C	Site
		Speed		30 rpm to 100 rpm	1.2 rpm	
		Time		15 min to 90 min	0.8 sec	
		Volume		500 ml to 1500 ml	8.0 ml	
10.2	Heat Transfer Press Machines	Temperature	WAGA/CM/031	100 °C to 200 °C	0.65 °C	Site
		Pressure		0 bar to 20 bar	0.061 bar	
		Time		5 sec to 30 sec	0.8 sec	

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location	
10.3	Metal detector	<b>Apparel Industry</b> (Nine point)	WAGA/CM/033	0.8 mm to 1.2 mm (ferrous)	*	Site	
		<b>Products / Packs (Convey belt)</b>					
		Dry Products (0 to 200 mm product height)	WAGA/CM/033	1.0 mm to 1.8 mm (ferrous)	*	Site	
				1.2 mm to 2.2 mm (non-ferrous aluminum)			
				1.5 mm to 3.0 mm (Stainless steel 316)			
		Wet Product (0 to 200 mm product height)	WAGA/CM/033	1.8 mm to 3.0 mm (ferrous)	*	Site	
				2.5 mm to 4.0 mm (non-ferrous aluminum)			
				3.5 mm to 5.0 mm (Stainless steel 316)			
		Aluminum Foil Pack (0 mm to 100 mm Product height)	WAGA/CM/033	1.0 mm to 4.0 mm (ferrous)	*	Site	
		Aluminum Foil Pack (100 mm to 200 mm Product height)		1.0 mm to 1.6 mm (ferrous)			
		<b>Free fall Vertical Packing Application</b>					
		Dry products (0 to 250 mm apparatus diameter)	WAGA/CM/033	1.0 mm to 1.8 mm (ferrous)	*	Site	
				1.2 mm to 2.0 mm (non-ferrous / aluminum)			
				1.5 mm to 2.5 mm (Stainless steel 316)			
		Wet / Frozen products and Metalized film packed products (0 to 250 mm aperture diameter)	WAGA/CM/033	1.5 mm to 2.5 mm (ferrous)	*	Site	
				2.0 mm to 3.2 mm (non-ferrous / aluminum)			
				2.5 mm to 4.0 mm (Stainless steel 316)			
		<b>Pipeline Application (Liquids, Slurries and Pastes)</b>					
		Wet Products (0 to 100 mm internal pipe diameter)	WAGA/CM/033	1.5 mm to 2.5 mm (ferrous)	*	Site	
				2.5 mm to 3.0 mm (non-ferrous / aluminum)	*		
				3.0 mm to 4.0 mm (Stainless steel 316)	*		

C.N. Gnan

Director / CEO

Sri Lanka Accreditation Board for Conformity Assessment