

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 15 January 2021
to 14 January 2024
Issued on 15 January 2021



ISO/ IEC 17025
CL 10-01

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: CL 010-01

Cassims International (Pvt) Ltd

255. Vauxhall Street
Colombo 02

Scope of Accreditation: Performing calibration on Colour temperature and Illuminance using PR 650 Spectroradiometer

The laboratory is accredited for the following calibrations.

SI No.	Type of Instrument	Calibration Performed	Calibration methods / Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
1.1	Visual Assessment Cabinet	Colour temperature	MM-CC-01 Rev 01:2018	Horizon (2100-2500) D75 (7300-7700) D65 (6300-6700) Inca A (2656-3056) TL 84 (3900-4300) CWF (3950-4350) U30 (2800-3200) U35 (3300-3700) UV	± 40 ± 45 ± 40 ± 35 ± 35 ± 35 ± 35 ± 35 *	Site
1.2		Illuminance	MM-CC-01 Rev 01:2018	Horizon 0-500 fCd D75 0-500 fCd D65 0-500 fCd Inca A 0-500 fCd TL 84 0-500 fCd CWF 0-500 Cd U30 0-500 fCd U35 0-500 fCd UV 0-10 fCd	± 6 ± 5 ± 4 ± 6 ± 10 ± 12 ± 10 ± 12 ± 1	Site

Note: fCd = femtoCandela

Director / CEO
Sri Lanka Accreditation Board for Conformity Assessment