



Valid from 11 March 2020
to 10 March 2023
Issued on 11 March 2020

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/ IEC 17025
TL 029-02

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: TL 029-02

Physical Testing Laboratory DPL Premier Gloves Limited Block B, BEPZ Walgama, Malwana

Scope of Accreditation: Performing Mechanical testing on Rubber Products as per the test methods appearing in this schedule

The laboratory is accredited for the following tests.

Sl No	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Rubber Products	Abrasion Resistance	EN388: 2016+A1 2018	Min 10 rubs Max 25000 rubs (level 1 to level 4)
		Blade Cut Resistance	EN388: 2016+A1 2018	Min 1.0 index Max 4.0 index (level 1 to level 3)
			EN ISO 13997: 1999	Min 2N Max 50N (level A,B,C,D,E and F)
		Tear Resistance	EN388: 2016+A1 2018	Min 5N Max 300N (level 1 to level 4)
		Puncture Resistance	EN388: 2016+A1 2018	Min 5N Max 300N (level 1 to level 4)
		Tensile Strength	ASTM D 412 : 2016	Min 10MPa Max 50 MPa

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment