



Valid from 17 May 2024
to 16 May 2028
Issued on 11 December 2024

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
CL 006-01

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 006-01

**Electruments International (Pvt) Ltd,
No. 371/C, Main Street,
Ganemulla.**

Scope of Accreditation: Performing Mass and Related Quantities, Length, Time and
Frequency & Thermal Calibrations and Measurement verification
of instruments as per the calibration methods appearing in the schedule.

The Laboratory is accredited for the following tests appear from page 02 to 05;

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution as applicable	Calibration Measurement Capability (CMC) (At 95% confidence Level, k=2) (±)	Location (Site/ In house)		
Thermal									
01	Calibration of Glass Thermometer	Temperature	ECAL/SOP/T05 Rev:01	0 to 200 °C	0.1 °C	0.19 °C	In house		
								0.2 °C	0.25 °C
								0.5 °C	0.4 °C
								1 °C	0.72 °C
02	Calibration of Digital Thermometer, Temperature Indicator with Sensor		Temperature	ECAL/SOP/T02 Rev:03	-30 to 50 °C	0.01 °C	0.62 °C	Site/ In house	
					<50 to 200 °C		0.16 °C		
					<200 to 400 °C		0.36 °C		
					-30 to 50 °C	0.1 °C	0.62 °C		
					<50 to 200 °C		0.17 °C		
					<200 to 400 °C		0.37 °C		
		-30 to 50 °C			1 °C	0.84 °C			
		<50 to 200 °C				0.65 °C			
		<200 to 400 °C				0.70 °C			
		-30 to 400 °C			2 °C	1.3 °C			
		5 °C	3 °C						
03	Calibration of Bottle Cooler, Refrigerator, Chiller	Temperature	ECAL/SOP/T03 Rev:01	(-10 to 20) °C	0.1 °C	0.9 °C	Site/ In house		
	Calibration of Freezer			(-30 to 5) °C					
	Calibration of Liquid Bath			(-10 to 250) °C					
	Calibration of Oven, Incubator, Dryer			(0 to 250) °C					
04	Calibration of Hygrometer (25°C and at RH 50%)	Temperature	ECAL/SOP/T04 Rev:06	@50 %			Site/ In house		
				RH 15°C - 25°C	0.1	0.8 °C			
				RH <25°C - 40°C		1.1 °C			
		Humidity		@25 °C					
				20 –35 RH %	0.1	1.2 (%)			
				55 – 70 RH %		2.0 (%)			
				70 – 80 RH %		2.2 (%)			
Length									
05	Calibration of Sample/ Circular cutter diameter	Length	ECAL/SOP/M02 Rev: 5	20 - 200 mm	0.02 mm	0.11 mm	Site/ In house		

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution as applicable	Calibration Measurement Capability (CMC) (At 95% confidence Level, k=2) (\pm)	Location (Site/ In house)
Mass and related quantities							
06	Pirspirometer Weight Calibration	Mass	ECAL/SOP/M07 Rev:07	0 - 5000 g	0.01 g	0.12 g	Site/ In house
07	Calibration of Analytical Balance, Precision Balance, Single pan, Top Loading Electronic Balance	Mass	ECAL/SOP/M01 Rev:08	(0-500) g	0.001 g	0.0030 g	
				(<500- 1000)g	0.01 g	0.050 g	
				(<1000- 10000)g	0.1 g	1.0 g	
				(<10000- 20000)g	1 g	1.5 g	
08	Calibration of Pressure Gauge, Pressure Transducers (Pneumatic)	Pressure	ECAL/SOP/M09 Rev:04	0 -1000 Pa (0 - 20) bar	1 Pa 0.01 bar	4.0 Pa 0.070 bar	
				0 -1000 Pa	1 Pa	4.0 Pa	
09	Calibration of Bursting Strength Machine Pressure Gauge (Pneumatic)	Pressure	ECAL/SOP/M09 Rev:04	(0 - 20) bar	0.01 bar	0.070 bar	
10	Force verification of Vertical Tensile Testing Machines and force measuring instruments with digital display or dial indicator	Force	ECAL/SOP/M10 Rev:07	(0 - 100) N (<300 - 1000) N		1.5 N 0.5%	
				(<1000 – 5000) N (0 – 100) N		5% 1.5 N	
				(<300 – 1000) N		0.5%	
Time and Frequency							
11	Calibration of Stop Watch/ Timers/ Machine Inbuild Timers (Observe method)	Time	ECAL/SOP/TI-01 Rev:03	30 sec	0.01 sec	1.2 sec	
				<30 sec to1800 sec	0.01 sec	1.2 sec	
				<30 sec to1800 sec	0.1 sec	1.2 sec	
				<30 sec to1800 sec	1 sec	1.4 sec	
				<1800 sec to 3600sec	1 sec	2.0 sec	
				<3600 sec to 32400 sec	1 sec	4.1 sec	
				<1 min to 540 min	1 min	35 sec	
12	Calibration of Stop Watch/ Timers/ Machine Inbuild Timers (Photo method)	Time	ECAL/SOP/TI-01 Rev:03	30 sec	0.01 sec	0.16 sec	
				<30 sec to1800 sec	1 sec	0.62 sec	
				<1800 sec to 3600 sec	1 sec	1.7 sec	
				<3600 sec to 32400 sec	1 sec	3.9 sec	

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution as applicable	Calibration Measurement Capability (CMC) (At 95% confidence Level, k=2) (±)	Location (Site/ In house)
13	Calibration of Stop Watch/ Timers/ Machine Inbuild Timers (Observe method)	Time	ECAL/SOP/TI-01 Rev:03	30 sec	0.01 sec	1.2 sec	Site/ In house
				<30 sec to 1800 sec	0.01 sec	1.2 sec	
				<30 sec to 1800 sec	0.1 sec	1.2 sec	
				<30 sec to 1800 sec	1 sec	1.4 sec	
				<1800 sec to 3600 sec	1 sec	2.0 sec	
				<3600 sec to 32400 sec	1 sec	4.1 sec	
				<1 min to 540 min	1 min	35 sec	
14	Calibration of Stop Watch/ Timers/ Machine Inbuild Timers (Photo method)	Time	ECAL/SOP/TI-01 Rev:03	30 sec	0.01 sec	0.16 sec	
				<30 sec to 1800 sec	1 sec	0.62 sec	
				<1800 sec to 3600 sec	1 sec	1.7 sec	
				<3600 sec to 32400 sec	1 sec	3.9 sec	
15	Calibration of Tachometer, Stroboscope (Optical Method Using Tachometer)	RPM	ECAL/SOP/M08 Rev:03	(30- 100) RPM	0.1 RPM	0.4 RPM	
				(<100 - 1000) RPM	0.1 RPM	0.25%	
				(<1000- 10000) RPM	1 RPM	0.10%	
				(<10000- 90000) RPM	1 RPM	0.06%	
16	Calibration of Inbuild RPM Device (Optical Method Using Tachometer)	RPM	ECAL/SOP/M08 Rev:03	(30- 100) RPM	0.1 RPM	0.4 RPM	
				(<100 - 1000) RPM	0.1 RPM	0.25%	
				(<1000- 10000) RPM	1 RPM	0.10%	
				(<10000- 90000) RPM	1 RPM	0.06%	
Measurement verification							
17	Measurement Verification of Temperature controlled Tumble dryer	Temperature	ECAL/SOP/T03 Rev:01	20°C to 30°C	0.1 °C	1 °C	Site/ In house
				<30°C to 90°C	0.1 °C	2.5 °C	
		RPM	ECAL/SOP/M08 Rev:03	(35- 60) RPM	0.1 RPM	0.50 RPM	
				Time	ECAL/SOP/TI-01 Rev:03	Cool down Time 10 min	
Heating Time up to 90 min	1 sec	2.2 sec					
18	Measurement Verification of Temperature controlled AATCC Washing Machine	Temperature	ECAL/SOP/T03 Rev:01	27°C - 41°C	0.1	0.5 °C	
				<41°C -95°C	0.1	1.1 °C	
		RPM	ECAL/SOP/M08 Rev:03	25 RPM- 100 RPM	0.1 RPM	0.6 RPM	
				(500 RPM-700 RPM)	0.1 RPM	0.40%	
		Time	ECAL/SOP/TI-01 Rev:03	Spin time (5 min)	1 sec	1.6 sec	
				Washing time (16 min)	1 sec	1.6 sec	
				Washing time (12 min)	1 sec	1.6 sec	
Washing time (8.5 min)	1 sec	1.6 sec					
19	Measurement Verification of Temperature controlled Standard washer	Temperature	ECAL/SOP/T03 Rev:01	27°C - 41°C	0.1 °C	0.5 °C	
				<41°C -80°C	0.1 °C	0.8 °C	
				<80°C -90°C	0.1 °C	1.0 °C	
		RPM	ECAL/SOP/M08 Rev:03	(40-100) RPM	0.1 RPM	0.80%	
				(500-1000) RPM	0.1 RPM	0.24%	
				(<1000-1150) RPM	0.1 RPM		
Time	ECAL/SOP/TI-01 Rev:03	Up to 3 hr	1 sec	3.0 sec			

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution as applicable	Calibration Measurement Capability Value (CMC) (At 95% confidence Level, k=2) (±)	Location (Site/ In house)
Measurement verification							
20	Calibration of Martindale Abrasion & Pilling machine	RPM	ECAL/SOP/ M08 Rev:03	(44.5 - 47.5) RPM	0.1 RPM	0.4 RPM	Site/ In house
		Mass	ECAL/SOP/ M07 Rev:07	Up to 3000 g	0.01 g	0.13 g	
21	Measurement Verification of Temperature controlled Front Loading Domestic Washing Machine	Temperature	ECAL/SOP/ T03 Rev:01	30°C -60°C	0.1	0.8 °C	
		RPM	ECAL/SOP/ M08 Rev:03	(30-100) RPM	1 RPM	0.7 RPM	
				(500-1000) RPM	1 RPM	2 RPM	
				(<1000 -1500) RPM	1 RPM	2.4 RPM	
Time	ECAL/SOP/ TI-01 Rev:03	Up to 02 hr	1 sec	2.3 sec			
22	Measurement Verification of Temperature controlled Durability washer	Temperature	ECAL/SOP/ T03 Rev:01	(50-90)°C	0.1	0.5 °C	
		RPM	ECAL/SOP/ M08 Rev:03	(500-3000) RPM	0.1 RPM	0.30%	
		Time	ECAL/SOP/ TI-01 Rev:03	up to 1 hr	1 sec	2.0 sec	
23	Measurement Verification of Temperature controlled Color Fastness Washing Machine	Temperature	ECAL/SOP/ T03 Rev:01	30°C -60°C	0.1	0.6 °C	
		RPM	ECAL/SOP/ M08 Rev:03	(30-50) RPM	0.1 RPM	0.40 rpm	
		Time	ECAL/SOP/ TI-01 Rev:03	up to 1 hr	1 sec	1.6 sec	

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment