



Valid from 31 May 2021  
To 30 May 2024  
Issued on 31 May 2021

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



## Schedule of Accreditation

Accreditation Scheme for Calibration Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 015-01

**Biomedical Engineering Laboratory,  
Wickramarachchi Institute of Speech and Hearing,  
J M Wickramarachchi & Company Private Limited  
No. 436, Delkanda,  
Nugegoda.**

**Scope of Accreditation:** : Performing calibration of pure tone audio meters as per  
acoustic method appearing in this Schedule.

The laboratory is accredited for the following tests as per given in the page 02 of 02.

SI	Parameter/ Measured Quantity/ Instrument or Gauge	Method of Calibration	Range	Readability / Resolution as applicable	Calibration Measurement Capability (Approximately at 95% Confidence Level)	Location
<b>Sound and Vibrations</b>						
01	Sound and Vibrations	ISO 389-1:2017 Calibration of sound pressure level for pure tones and supra aural headphones  IEC 60645-1:2017 Equipment for pure tone and speech audiometry	-5 – 120dB	5dB	±3	Laboratory
		ISO 389-3:2016 Calibration of vibratory force level for pure tones and bone vibration  IEC 60645-1:2017 Equipment for pure tone and speech audiometry	-5 – 70dB			

Director/CEO  
Sri Lanka Accreditation Board for Conformity Assessment