

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 30 June 2020
to 27 October 2021
Issued on 30 June 2020



Schedule of Accreditation

Accreditation Scheme for Testing/Calibration Laboratories Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 007-01

Lanka Calibration Services (Pvt) Ltd

**No. 27/14A
Rosmead Place
Colombo 07.**

Scope of Accreditation: Performing Thermal as per the calibration methods given in the schedule of accreditation.

The laboratory is accredited for the calibration as given below

SI No	Type of instrument	Calibration performed	Calibration methods / Measurement procedure	Range of calibration	CMC values	Location
01	Humidity Data Loggers with probe using Humidity Chamber	Humidity & Temperature	LCS/TM/24: Rev. 0	20 - 90 % RH	3 %	LCS Lab 2

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment