



Valid from 11 August 2021  
to 11 December 2022  
Issued on 11 August 2021

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



ISO/IEC 17025  
CL 017-01

## Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 017 -01

**Professional MET Consultancy Services (Pvt) Ltd**

**No: 04A, Ohlums Place, Borella**

**Colombo 08.**

**Scope of Accreditation:** Performing Thermal calibrations and Mechanical calibration (Pressure and Length) as per the calibration methods appearing in the schedule.

The Laboratory is accredited for the following calibrations.

Sl NO	Type of Instrument	Calibration Performed	Calibration Methods/ Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
<b>Mechanical Calibration</b>						
01	Feeler Gauge	Mechanical Calibration	PROMET/DI/005	0 to 1 mm	0.0009 mm	In-house

SI NO	Type of Instrument	Calibration Performed	Calibration Methods/ Measurement Procedure	Range of Calibration	Calibration Measurement Capability	Location
02	Test Sieve	Mechanical Calibration	PROMET/DI/006	1 to 200 mm	0.015 mm	In-house
03	Thickness Gauge	Mechanical Calibration	PROMET/DI/007 Rev-01	0 to 20mm /0.01 mm	0.006 mm	In-house
				0 to 20mm /0.001 mm	0.0006 mm	In-house
04	Pressure Gauge	Mechanical Calibration	PROMET/PR/001 Rev-03	0 to 30 bar / 0.1 bar	0.14 bar	In-house/on site
<b>Routine Calibrations of Calibrations for the Reference Standards</b>						
05	Calibration of Micrometer	Mechanical Calibration	PROMET/DI/004 Rev-01	0 to 25 mm/ 0.001 mm	0.0006 mm	In-house
				0 to 1000 mm/ 0.01 mm	0.006 mm	
06	Feeler Gauge	Mechanical Calibration	PROMET/DI/005	0 to 1 mm	0.0009 mm	In-house

Director/CEO

Sri Lanka Accreditation Board for Conformity Assessment