



## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 074 – 01

Water Quality Monitoring Laboratory Central Environmental Authority "Parisara Piyasa", No 10, Densil Kobbekaduwa Mawatha Baththaramulla.

**Scope of Accreditation:** Performing Chemical Testing on Water and Waste Water. The laboratory is accredited for the following tests.

SI	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Water and Waste Water	Total Suspended Solid	APHA 2540 D 2017 23 <sup>rd</sup> Edition	1 ppm
		Nitrate	APHA 4500 NO <sub>3</sub> - E 2017 23 <sup>rd</sup> Edition	0.5 ppm
		Total Phosphorous	APHA 4500 P-B & E 2017 23 <sup>rd</sup> Edition	0.02 ppm
		Total Dissolved Solids	APHA 2540 C 2017 23 <sup>rd</sup> Edition	28 ppm

Director /CEO Sri Lanka Accreditation Board for Conformity Assessment