



Valid from 30 June 2020  
to 29 June 2023  
Issued on 30 June 2020

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



ISO/ IEC 17025  
TL 019-01

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 019-01

**Astron Ltd**  
**No. 688, Galle Road**  
**Ratmalana.**

**Scope of Accreditation:** Performing Biological Testing on Pharmaceutical Liquid Preparations, Water, Milk Powder & Soya based products as per SLS and USP methods.

The laboratory is accredited for the tests given in page 02.

Sl	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed (eg: xxx: 2016)	Range of testing/ Limits of detection
<b>Biological Testing</b>				
1	Water	Coliform/100mL	SLS 614 : 2013 SLS 1461 - Part 1 Section 3	Nil to 1800 MPN
		<i>E.Coli</i> /100mL	SLS 614 : 2013 SLS 1461 - Part 1 Section 3	Nil to 1800 MPN
2	Milk Powder & Soya based products	Aerobic Plate Count/g or mL	SLS 731 : 2008 SLS 516 - Part 1 Section 1  USP - 44 NF - 39 - 2021 Section 2021	10 to $2 \times 10^4$ cfu/g or mL  10 to $3 \times 10^3$ cfu/g or mL
		Coliform/g or mL	SLS 731 : 2008 SLS 516 – Part 3 Section 1	Nil to 110 MPN
		<i>Staphylococcus aureus</i> / g or mL	USP - 38 NF - 33 : 2015 Section 2022	Present or Absent
		<i>E. Coli</i> / g or mL	SLS 731 : 2008 SLS 516 - Part 12 : 2013  USP - 44 NF - 39 : 2021 Section 2022	Nil to 110 MPN  Present or Absent
3	Pharmaceutical Liquid Preparations	Aerobic Plate Count / g or mL	USP - 44 NF - 39 - 2021: Section 2021	10 to $3 \times 10^3$ cfu/g or mL
		<i>E. Coli</i> / g or mL	USP - 44 NF - 39 - 2021: Section 2022	Present or Absent
		<i>Staphylococcus aureus</i> / g or mL	USP - 38 NF - 33 - 2015: Section 2022	Present or Absent

Director/CEO  
Sri Lanka Accreditation Board for Conformity Assessment