



Valid from 19 January 2023
to 29 June 2026
Issued on 19 January 2024

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
TL 019-02

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 019-02

Astron Ltd
No. 688, Galle Road
Ratmalana.

Scope of Accreditation: Performing Chemical testing on Asartan tablet and Mechanical testing on Pharmaceutical Tablet Preparations (Metformin ER Tablets, Clopidogrel 75 mg Tablets, Meben tablets, Elevate tablets) as per the Test Methods appearing in this schedule.

The Laboratory is accredited for the following tests appear on page 02;

SI	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed (eg: xxx:2016)	Range of testing/ Limits of detection	
Chemical Testing					
1	Asartan 50 mg tabs (Losartan Potassium USP 50mg Tablets)	Assay for Losartan Potassium	USP 2023 Losartan Potassium Tablets Monograph	95-105% of label claim	
Mechanical Testing					
1	Pharmaceutical Tablet Preparations (For all tablets mentioned below)	Weight variation (mg)	BP Section 2.9.5:2020	80mg–250mg: $\pm 7.5\%$ 250mg or more: $\pm 5\%$ Tab wts range: 145 - 900mg	
2		Friability (%) (*This test is valid only for uncoated tablets)	BP Appendix XVII G Section 2.9.7: 2020	0 - 1%	
3		Disintegration (minutes)	BP Section 2.9.1:2020	Uncoated tabs: 0-15 min Coated tabs: 0 – 30 min	
4		1. Metformin ER Tablets 2. Clopidogrel 75mg Tablets	Hardness (kg)	AST/QO/QC-TM/SOP5 Issue Date: 01/10/2019 Issue No.: 01 Revision No: 00	0.3– 50 kg
5		3. Meben Tablets 4. Elevate Tablets	Thickness (mm)	AST/QO/QC/SOP3 Issue Date: 01/06/2021 Revision Date: 01/01/2022 Issue No.: 01 Revision No: 01	0.1 – 12 mm
6			Diameter (mm)	AST/QO/QC/SOP3 Issue Date: 01/06/2021 Issue No.: 01 Revision No: 01	0.1 – 26 mm

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment