



Valid from 07 April 2021  
To 06 April 2024  
Issued on 06 April 2024

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



ISO/ IEC 17025  
TL 020-02

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 020-02

**Chemical Laboratory**  
**MicroChem Laboratories (Pvt) Ltd**  
**No. 134, Dehiwala Road**  
**Bellanwila, Borelasgamuwa.**

**Scope of Accreditation:** Performing Chemical Testing on Water & Waste Water as per APHA test methods.

The laboratory is accredited for the tests given below.

Sl	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Drinking Water Potable water Waste water Swimming pool water	pH	APHA 4500 - H <sup>+</sup> B 23 <sup>rd</sup> edition:2017	0 - 14
		Total Hardness, mg/l	APHA 2340 - C 23 <sup>rd</sup> edition:2017	5 (Minimum)
		Alkalinity, mg/l	APHA 2320 B 23 <sup>rd</sup> edition:2017	3 (Minimum)
		Chloride, mg/l	APHA 4500 B 23 <sup>rd</sup> edition:2017	4 (Minimum)

Director /CEO  
Sri Lanka Accreditation Board for Conformity Assessment