



Valid from 22 August 2024
to 29 October 2025
Issued on 22 August 2024

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/ IEC 17025
TL 006- 01

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: TL 006-01

Lindel Industrial Laboratories Limited

Pattiwila Road, Sapugaskanda, Makola.

Scope of Accreditation: Performing Chemical Testing on Waste Water, Fuel oil & Fertilizer as per SLS, ASTM & APHA Methods appearing in this schedule.

The laboratory is accredited for the following tests as per given in the page 02.

Sl	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Waste Water	Ammonical Nitrogen	APHA 4500 – NH ₃ B & C 24 th Edition	5-100 ppm
02	Fuel oil (Crude petroleum, petroleum products / mixtures of petroleum & non pet liquid products)	Density	ASTM D 1298: 2021	700 – 1000 Kg/m ³ at 15 ^o C
03	Fuel oil (Distillate and residual fuels, gas turbine fuels, crude oils, lubricating oils, waxes & other pet products)	Ash content	ASTM D 482:2021	0.01% - 0.18%
04	Fuel oil (Petroleum products / mixtures of petroleum & non pet liquid products)	Flash point	ASTM D 93:2021	40 ^o C – 150 ^o C
05	Fertilizer	Total Nitrogen as N (chemical, inorganic & organic)	SLS 645: Part 1:2009	5% - 50%
06		Total Phosphate (as P ₂ O ₅)	SLS 645: Part 5: 1985	0.3% – 70%
07		Moisture	SLS 645: Part 2: 1984	0.1% - 25%

.....

Chairman
Sri Lanka Accreditation Board for Conformity Assessment