

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid until 05 June 2016  
Issued on 06 June 2013



ISO 15189  
ML 018-01

## Schedule of Accreditation

Accreditation Scheme for Medical/Clinical Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: ML 018-01

**The Royal Hospital Laboratory  
Royal Hospital (Pvt) Ltd.,  
No. 62, W A De Silva Mawatha,  
Colombo 06.**

**Scope of Accreditation:** Performing Medical Tests under the fields of Clinical Biochemistry, Haematology and Clinical Pathology

The laboratory is accredited for the following tests. (Please see page 02 onwards for details)

	<b>Field of Testing</b>	<b>Test</b>	<b>Test Method</b>	<b>Test Instrument</b>	<b>Analytical Range</b>	<b>MU (±)</b>
<b>01</b>	<b>Clinical Biochemistry</b>	Plasma Glucose Fasting plasma glucose Postprandial plasma glucose	Hexokinase	Cobas c111 Biochemistry analyzer	1.98-720 mg/dL	2.6 mg/ dL
		Serum Urea	Urease		3-240 mg/dL	2.9 mg/ dL
		Serum Creatinine	Enzymatic		0.06-30.5 mg/dL	0.09 mg/ dL
		Serum Cholesterol	Enzymatic CE-CHOD- POD		9.7-800 mg/dL	2.8 mg/ dL
		Serum Triglycerides	Enzymatic GPO-POD		8.85-885 mg/dL	4.5 mg/ dL
		Serum SGPT/ ALT	IFCC Kinetic		2-700 U/ L	1.7 U/L
		Serum SGOT/ AST	IFCC Kinetic		2-700 U/ L	2.4 U/L
<b>02</b>	<b>Clinical Pathology</b>	Urine Full Report	Manual Strip Method	Dry Chemistry Strips and Microscope	Presence or absence and semi quantitative	-
		Stool Full Report (AOC)	Manual wet smear	Microscopy	Presence or absence	-
		Urine HCG	Manual	Reagent Strips	Negative / Positive	-
<b>03</b>	<b>Haematology</b>	ESR	Manual	Westegran method	1-100 mm	-
		RF	Manual agglutination	Manual	Negative or Positive	-

Director /CEO  
Sri Lanka Accreditation Board for Conformity Assessment