



Valid from 01 July 2019  
to 30 June 2022  
Issued on 01 July 2019

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



ISO/ IEC 17025  
TL 056-02

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment  
Accreditation Number: TL 056-02

### IGBU QC Laboratory Ansell Lanka (Pvt) Ltd Biyagama EPZ, Biyagama

**Scope of Accreditation:** Performing Mechanical testing on Rubber & Textile gloves as per BS EN and ISO methods

The laboratory is accredited for the following tests.

SI No.	Product(s) / Material of test	Specific tests performed	Test method / Standard against which tests are performed	Range of testing / Limits of detection
1.1	Rubber & Rubber products (PPE; Gloves & Sleeves)  Textile & related products (PPE; Gloves & Sleeves)	EN Abrasion resistance	BS EN 388:2016, Clause 6.1	Min 100 Rev Max 30000 Rev (Level 1 – Level 4) Limit of detection: 100 Rev
1.2		EN Blade cut resistance	BS EN 388:2016, Clause 6.2	Up to Index 20.0 (Up to Level 5) Limit of detection: 0.1 Index
1.3		EN ISO Cut resistance	EN ISO 13997:1999 (Product standard) BS EN 388:2016, Clause 6.3	1N – 40N (Level A – Level F) Limit of detection: 0.1 N
1.4		EN Tear resistance	BS EN 388:2016, Clause 6.4	Min 10 N Max 200 N (Up to Level 4) Limit of detection: 0.1 N
1.5		EN Puncture resistance	BS EN 388:2016, Clause 6.5	Min 10 N Max 200 N (Up to Level 4) Limit of detection: 0.1 N