



Valid from 31 May 2024
to 30 May 2028
Issued on 11 September 2028

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/ IEC 17025
TL 060-02

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: TL 060-02

Refrigerator Testing Laboratory
Energy & Environment Services Department
National Engineering Research & Development Centre
2P, 17B, Industrial Estate, Ekala Ja-Ela

Scope of Accreditation: Performing Electrical testing on refrigerators as per SLS 1690

The laboratory is accredited for the following tests.

SI No.	Product(s) / Material of test	Specific tests performed	Test method / Standard against which tests are performed	Range of testing / Limits of detection
01	Household refrigerators	Energy use test	SLS 1690:2020 (Clause 6.4)	1 – 1000 Wh
		Temperature performance test	SLS 1690:2020 (Clause 6.3)	-18 – 15 °C

Chairman
Sri Lanka Accreditation Board for Conformity Assessment