As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 11 August 2024 to 10 August 2028 Issued on 17July 2024



ISO/ IEC 17025 CL 019 - 01

## Schedule of Accreditation

Accreditation Scheme for Calibration Laboratories Sri Lanka Accreditation Board for Conformity

AssessmentAccreditation Number: CL 019-01

## Utracon Calibration Laboratory No:100/1, Pahala Imbulgoda, Imbulgoda

**Scope of Accreditation:** Performing Mechanical calibration (Force) as per the calibration method appearing in the schedule

The laboratory is accredited for the following calibrations;

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readabil ity/ Resoluti on as applicab le	Calibration Measurement Capability	Locati on (Site/ In house)
01	Hydraulic Jacks (100 kN – 1000 kN)	Measurement of Static Force in Compression	ISO 7500-1:2018	100 kN – 1000 kN	0.1 kN	1.5 %	In
02	Hydraulic Jacks (200 kN – 5000 kN)	Measurement of Static Force in Compression	ISO 7500-1:2018	200 kN – 5000 kN	0.i kN	2.2 %	House

Chairman

Sri Lanka Accreditation Board for Conformity Assessment