



Valid from 15 November 2022  
to 14 November 2025  
Issued on 15 November 2022

As an accredited laboratory, this laboratory is entitled to  
use the following accreditation symbol.



ISO/IEC 17025  
TL 093-01

## Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 093-01

**Lalanka Laboratory Services (Pvt) Ltd,  
No. 13/3, Sri Dharmarama Road,  
Ratmalana.**

**Scope of Accreditation:** Performing Chemical testing on water and waste water as per the APHA methods.

The Laboratory is accredited for the following tests appear on page 02;

<b>SI NO</b>	<b>Product(s) / Material of test</b>	<b>Specific tests performed</b>	<b>Test Method/ Standard against which tests are performed</b>	<b>Range of testing/ Limits of detection</b>
01	Water	pH	APHA 4500 H <sup>+</sup> B :23 <sup>rd</sup> Ed	4.0 – 9.0
		Total Alkalinity	APHA 2030 B: 23 <sup>rd</sup> Ed	2 – 1000mg/L as CaCO <sub>3</sub>
		Total Hardness	APHA 2340 C :23 <sup>rd</sup> Ed	2 – 600 mg/L as CaCO <sub>3</sub>
		Electrical Conductivity	APHA 2510 B :23 <sup>rd</sup> Ed	1 – 1500 µS/cm
		Total Dissolved Solids	APHA 2540 C :23 <sup>rd</sup> Ed	4 – 1200 mg/L 1201 2000 mg/L
02	Waste Water	pH	APHA 4500 H <sup>+</sup> B :23 <sup>rd</sup> Ed	4.0 – 9.0
		Total Suspended solids	APHA 2540 D: 23 <sup>rd</sup> Ed	2 – 100 mg/L 101 600 mg/L

Director/CEO  
Sri Lanka Accreditation Board for Conformity Assessment