

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 01 February 2023
to 31 January 2026
Issued on 27 February 2023



ISO 15189
ML 039 - 01

Schedule of Accreditation

Accreditation Scheme for Medical/Clinical Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: ML 039-01

Medihelp Hospitals Laboratory (Pvt) Ltd
No 172/3, Panadura Road
Horana

Scope of Accreditation: Performing Medical testing under the fields of Clinical Biochemistry / Chemical Pathology, Clinical Pathology, Haematology

The laboratory is accredited for the following tests.

| Sl | Field of Testing | Test | Test Method | Test Instrument | Linearity Range |
|----|--|--|--|-----------------|----------------------|
| 01 | Clinical Biochemistry/ Chemical Pathology | Glucose (Plasma) | GOD – PAP (Glucose oxidase-phenol and 4 aminophenazone) | BS 800M | 10mg/dL to 500 mg/dL |
| | | Cholesterol (Serum) | CHOD – PAP (cholesterol oxidase/peroxidase aminophenazone) | | 1mg/dL to 500mg/dL |
| | | SGOT (serum glutamic-oxaloacetic transaminase) (Serum) | IFCC (International Federation of Clinical Chemistry) | | up to 260 U/L |
| | | SGPT (Serum glutamic pyruvic transaminase) (Serum) | | | up to 260 U/L |

| Sl | Field of Testing | Test | Test Method | Test Instrument | Linearity Range | |
|-----------|--|---|--|-----------------------------------|-----------------------------------|----------|
| 01 | Clinical Biochemistry/ Chemical Pathology | GGT (gamma-glutamyl transferase) (Serum) | IFCC (International Federation of Clinical Chemistry) | BS 800M | 4 – 650 U/L | |
| | | TSH (Thyroid Stimulating Hormone) (Serum) | Electrochemiluminescence | Cobas e 411 | 0.005 – 100.0 μIU/mL | |
| 02 | Clinical Pathology | Urine HCG (Urine) | Strip Method | Bio panda HCG Rapid Test Strip | +/- | |
| 03 | Haematology | FBC (Whole Blood) | | | | |
| | | Haemoglobin | Colorimetric | Mindray 5380 | 0 – 250 g/L | |
| | | WBC | Electrical Impedance | | 0.00 – 99.99 * 10 ⁹ /L | |
| | | RBC | | | 0.00 – 8.00 * 10 ¹² /L | |
| | | Platelet | | | 0 – 1000 * 10 ⁹ /L | |
| | | HCT | | | Calculated | 0 – 67 % |
| | | MCV | Computed from PCV & RBC (fl) | | | |
| | | MCH | Computed from Hb & RBC (pg) | | | |
| | | MCHC | Computed from Hb & PCV (g/dL) | | | |
| | | RDW | - | | | |
| | | Differential count (Whole Blood) | | | | |
| | | Neutrophils | Laser scattering | Mindray 5380 | - | |
| | | Lymphocytes | | | | |
| | | Monocytes | | | | |
| | | Eosinophils | | | | |
| Basophils | | | | | | |

Approved Sample Collection Centers

| No | Sample Collection Centre | Address | Matrix of the sample |
|-----|----------------------------------|--|----------------------|
| 1. | Medihelp Hospital - Horana | No.172/3, Panadura Road, Horana | Blood/Urine/Stool |
| 2. | Medihelp Hospitals - Piliyandala | No.100, Colombo Road, Piliyandala | Blood/Urine/Stool |
| 3. | Medihelp Hospitals - Moratuwa | No.246, Galle Road, Moratuwa | Blood/Urine/Stool |
| 4. | Medihelp Hospitals - Beruwala | No.173, Galle Road, Beruwala | Blood/Urine/Stool |
| 5. | Medihelp Hospitals - Kalutara | No.558, Galle Road, Kaluthara South | Blood/Urine/Stool |
| 6. | Medihelp Hospitals - Homagama | No.69, Highlevel Road, Hoamagama | Blood/Urine/Stool |
| 7. | Medihelp Bulathsinhala | No.201/1, Opposite to Hospital, Bulathsinghala | Blood/Urine/Stool |
| 8. | Medihelp Ingiriya | Ingiriya Medihelp, Bandaranayake Mawatha, Ingiriya | Blood/Urine/Stool |
| 9. | Medihelp Wethara | No.8/2, Rilawala, Wethara, Polgasowita | Blood/Urine/Stool |
| 10. | Medihelp Kesbewa | No.130, Horana Road, Kesbewa | Blood/Urine/Stool |
| 11. | Millaniya Collecting Center | No. 232, Horana Road, Millaniya | Blood/Urine/Stool |
| 12. | Eduragala Collecting Center | No.8/A, Eduragala, Ingiriya | Blood/Urine/Stool |
| 13. | Medihelp Hospitals -Athurugiriya | 175/3/D, Pore, Athurugiriya | Blood/Urine/Stool |
| 14. | Medihelp Hospitals - Kelaniya | 625, Bulugaha Junction, Kandy Road, Kelaniya | Blood/Urine/Stool |

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment
Colombo, Sri Lanka