



As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.

Valid from 21 November 2025
to 20 November 2029
Issued on 24 November 2025



ISO 15189
ML 039-01

Schedule of Accreditation

Accreditation Scheme for Medical/Clinical Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: ML 039-01

Medihelp Hospitals (Pvt) Ltd, Laboratory, Horana

Scope of Accreditation: Performing Medical testing under the fields of Chemical Pathology/ Clinical Biochemistry and Haematology

The laboratory is accredited for the following tests.

| Sl | Field of Testing | Test | Test Method | Test Instrument | Linearity Range |
|----|----------------------------------------------|----------------------------------------|-----------------------------------------------------------|-----------------|------------------|
| 01 | Chemical Pathology/ Clinical Biochemistry | Plasma Glucose | GOD-PAP (Glucose oxidase-Phenol and 4 aminophenazone) | Mindray BS 800 | 10 - 500 mg/dL |
| | | Serum Cholesterol | CHOD-PAP (cholesterol oxidase/peroxidase aminophena-Zone) | | 1 - 500 mg/dL |
| | | Serum AST (Aspartate aminotransferase) | IFCC (International Federation of clinical chemistry) | | 1 - 260 U/L |
| | | Serum ALT (Alanine Aminotransferase) | IFCC (International Federation of clinical chemistry) | | 1- 260 U/L |
| | | Serum GGT (gamma-glutamyl transferase) | IFCC (International Federation of clinical chemistry) | | 4 - 650 U/L |
| | | Serum Creatinine | enzymatic | | 0.14-13.56 mg/dL |
| | | Serum Urea | Urease-glutamate Dehydrogenase | | 0.9-40 mg/dL |

| Sl | Field of Testing | Test | Test Method | Test Instrument | Linearity Range | |
|----|----------------------------------------------|-----------------------------------------|--------------------------------|-----------------|----------------------------------|--------|
| 01 | Chemical Pathology/ Clinical Biochemistry | Serum TSH (Thyroid Stimulating Hormone) | Electrochemiluminescence | Cobas e411 | 0.005 - 100.0 μ U/mL | |
| 02 | Clinical Pathology | Urine HCG | FIA (Fluorescence immunoassay) | Strip Method | Positive or Negative | |
| 03 | Haematology | Full Blood Count | | | | |
| | | Haemoglobin | Colorimetric | Mindray 5380 | 0 - 2.5 g/dl | |
| | | WBC | Electrical Impedance | | 0.00 - 99.99x10 ⁹ /L | |
| | | RBC | | | 0.00 - 8.00x 10 ¹² /L | |
| | | Platelet | | | 0-1000x10 ⁹ /L | |
| | | HCT | | | Calculated | 0-67 % |
| | | MCV | Computed from PCV & RBC (fl) | | | |
| | | MCH | Computed from HB & RBC (pg) | | | |
| | | MCHC | Computed from HB & PCV (g/dL) | | | |
| | | RDW | | | | |
| | | Differential Count | Laser scattering | | | |

Approved Sample Collection Centers

| No | Sample Collection Centre | Address | Matrix of the sample |
|-----|-------------------------------------|----------------------------------------------------------|----------------------|
| 1. | Medihelp Hospital - Horana | No.172/3, Panadura Road, Horana | Blood/Urine/Stool |
| 2. | Medihelp Hospitals - Piliyandala | No.100, Colombo Road, Piliyandala | Blood/Urine/Stool |
| 3. | Medihelp Hospitals - Moratuwa | No.246, Galle Road, Moratuwa | Blood/Urine/Stool |
| 4. | Medihelp Hospitals - Beruwala | No.173, Galle Road, Beruwala | Blood/Urine/Stool |
| 5. | Medihelp Hospitals - Kalutara | No.558, Galle Road, Kaluthara South | Blood/Urine/Stool |
| 6. | Medihelp Hospitals - Homagama | No.69, Highlevel Road, Hoamagama | Blood/Urine/Stool |
| 7. | Medihelp Bulathsinhala | No.201/1, Opposite to Hospital, Bulathsinghala | Blood/Urine/Stool |
| 8. | Medihelp Ingiriya | Ingiriya Medihelp, Bandaranayake Mawatha, Ingiriya | Blood/Urine/Stool |
| 9. | Medihelp Wethara | No.8/2, Rilawala, Wethara, Polgasowita | Blood/Urine/Stool |
| 10. | Medihelp Kesbewa | No.130, Horana Road, Kesbewa | Blood/Urine/Stool |
| 11. | Medihelp Hospitals -Athurugiriya | 175/3/D, Pore, Athurugiriya | Blood/Urine/Stool |
| 12. | Medihelp Hospitals - Kelaniya | 625, Bulugaha Junction, Kandy Road, Kelaniya | Blood/Urine/Stool |
| 13. | Medihelp Hospital – Wellness Center | Rev. Baddegama Wimalawansa Mawatha, Colombo 007 | Blood/Urine/Stool |

Acting Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment
Colombo, Sri Lanka