

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 11 December 2024  
to 16 May 2026  
Issued on 11 December 2024



ISO/IEC 17025  
CL 021-01

## Schedule of Accreditation

Accreditation Scheme for Testing Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment  
Accreditation Number: CL 021-01

**Bureau Veritas Consumer Products Services Lanka (Pvt) Ltd**  
No.270, Dehiwala Road, Bellanwila

**Scope of Accreditation:** Performing Pressure calibrations as per in-house methods

The laboratory is accredited for the following calibrations.

SI No	Type of instrument	Calibration performed	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution on as applicable	Calibration Measurement Capability	Location
<b>Pressure</b>							
01	Pneumatic Pressure Gauge	Direct Comparison	CAL/SOP/MM/004	0 bar to 60 bar	Class 1.0 or lower	0.062 bar	Site/In-House
02	Hydraulic Pressure Gauge	Direct Comparison	CAL/SOP/MM/004	0 bar to 1000 bar	Class 1.0 or lower	1.2 bar	Site/In-House

Director / CEO  
Sri Lanka Accreditation Board for Conformity Assessment