

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 08 June 2023  
to 07 June 2026  
Issued on 15 July 2023



ISO 15189  
ML 042-01

## Schedule of Accreditation

Accreditation Scheme for Medical/Clinical Laboratories  
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: ML 042-01

**Asiri Laboratories, Asiri Hospital Kandy**  
**No. 907, Peradeniya Road, Kandy**

**Scope of Accreditation:** Performing Medical/Clinical testing under the fields of Clinical Biochemistry/Chemical Pathology, Clinical Pathology and Haematology

The laboratory is accredited for the following tests.

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range
01	Clinical Biochemistry / Chemical Pathology	Albumin (Serum)	BCG	Beckman Coulter AU480	1.1 -70 g/L
		SGPT (ALT) (Serum)	IFCC		1.6-800 U/L
		SGOT (AST) (Serum)			1.6 -800 U/L
		Cholesterol (Serum)			1.6 -800 U/L
		Creatinine (Serum)	CHOD-PAP		3-750 mg/dL
		Plasma Glucose	Enzymatic		0.05-30 mg/dL
			GOD-PAP		1.0 - 400 mg/dL

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range	
01	Clinical Biochemistry / Chemical Pathology	Triglycerides (Serum)	GPO-PAP	Beckman Coulter AU480	2-1000 mg/dL	
		Total protein (Serum)	BIURET		4.6-150 g/L	
		Urea (Serum)	UREASE		2.5 -300 mg/dL	
02	Clinical Pathology	<b>Urine Full Report</b>				
		Colour	Manual	-	-	
		Appearance				
		Specific Gravity	Strip method	Abbott-Urocolor 10 Strip	1.012-1.030	
		pH			5.0-9.0	
		Protein			Nil/Trace/+/++/+++ / +++++	
		Glucose			Nil/+/++/+++ / +++++	
		Ketone bodies			Nil/Trace/ +	
		Bilirubin			Nil/Trace/ +	
		Urobilinogen			Normal amounts / +	
		Nitrate			Nil/Trace/ +	
		Pus cells	Microscopic	Olympus CX43	Counted Cells	
		Red cells			Counted Cells	
		Epithelial cells			+ / ++	
		Casts			Nil/ (casts type) + a few/ +	
Crystals	Nil/ (casts type) + a few/ +					

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range	
03	Haematalogy	<b>Full Blood Count</b>				
		RBC	DC detection method	Sysmex XS 1000i	0 - 8 X 10 <sup>6</sup> /μL	
		WBC	Flow cytometric method		Within ± 3% or ± 0.3 X 10 <sup>3</sup> / μL t 100 X 10 <sup>3</sup> /μL	
			Differential Count		Neutrophils - 0-99.9% Lymphocytes- 0 -99.9% Monocytes- 0-99.9% Eosinophils – 0-99.9% Basophils- 0-99.9%	
		HGB	SLS Hemoglobin method		0.0 – 25.0 g/dL	
		PLT	DC detection method		0.0 - 5000 X 10 <sup>3</sup> /μL	
		HCT	Calculated		0.0 - 60.0 HCT %	
		MCH				
		MCV				
		MCHC				
		RDW-CV				
		ESR Whole Blood (Citrate/ EDTA)	Westergern Method			2-120 mm

## Approved sample collection centres

SI no	Name of the sample collection center	Address	Matrix of the sample
01	Asiri Laboratory - Kandy	No 82, William Gopallawa Mawatha, Kandy	Plain, EDTA, Urine, Stool
02	Asiri Collection Centre - Bogambara	No 46/1, Keppetipola Road, Kandy	Plain, EDTA, Urine, Stool
03	Asiri Collection Centre - Gampola	No 36C, T.B Panabokke Mawatha, Gampola	Plain, EDTA, Urine, Stool
04	Asiri Collection Centre - Katugasthota	No 206A, Katugasthota Rd, Kandy	Plain, EDTA, Urine, Stool
05	Asiri Collection Centre - Kegalle	No 325, Main street, Kegalle	Plain, EDTA, Urine, Stool
06	Asiri Collection Centre - Mawanella	No 209, Colombo Road, Mawanella	Plain, EDTA, Urine, Stool
07	Asiri Collection Centre - Nawalapitiya	No 119, Kotmale Road, Nawalapitiya	Plain, EDTA, Urine, Stool
08	Asiri Collection Centre - Peradeniya	No 116, Peradeniya Road, Kandy	Plain, EDTA, Urine, Stool
09	Asiri Collection Centre - Pilimathalawa	No 197/1, Daulagala Road, Pilimathalawa	Plain, EDTA, Urine, Stool
10	Asiri Collection Centre - Thennekumbura	No 124, Thennekumbura, Kandy	Plain, EDTA, Urine, Stool
11	Asiri Collection Centre- Geliyoa	No 62, Kandy Rd, Geliyoa	Plain, EDTA, Urine, Stool
12	Asiri Collection Centre- Akurana	No 237/2/B, Matale Road , Akurana	Plain, EDTA, Urine, Stool
13	Asiri Collection Centre - Wattegama	No 45, Mathale Road, Waththegama	Plain, EDTA, Urine, Stool
14	Asiri Collection Centre - Dambulla	No 712 , Anuradhapura Road, Dambulla	Plain, EDTA, Urine, Stool
15	Asiri Laboratories - Matale	No 91/A, Kandy Road, Matale	Plain, EDTA, Urine, Stool

Director/CEO  
Sri Lanka Accreditation Board for Conformity Assessment