

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 05 December 2025
to 04 December 2029
Issued on 23 February 2026



ISO 15189
ML 042-01

Schedule of Accreditation

Accreditation Scheme for Medical/Clinical Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: ML 042-01

**Asiri Laboratories, Asiri Hospital Kandy
No. 907, Peradeniya Road, Kandy**

Scope of Accreditation: Performing Medical/Clinical testing under the fields of Clinical Biochemistry/Chemical Pathology, Clinical Pathology and Haematology

The laboratory is accredited for the following tests.

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range
01	Clinical Biochemistry / Chemical Pathology	Albumin (Serum)	BCG	Beckman BC AU 480	1.43 -70 g/L
		SGPT (ALT) (Serum)	IFCC		1.6 – 800 U/L
		SGOT (AST) (Serum)			1.6 - 800 U/L
		Cholesterol (Serum)	CHOD-PAP		3 - 750 mg/dL
		Creatinine (Serum)	Enzymatic		0.05 - 30 mg/dL
		Plasma Glucose	GOD-PAP		1.0 - 400 mg/dL

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range		
01	Clinical Biochemistry / Chemical Pathology	Triglycerides (Serum)	GPO-PAP	Beckman BC AU 480	2-1000 mg/dL		
		Total protein (Serum)	BIURET		4.6-150 g/L		
		Urea (Serum)	UREASE		2.5 -300 mg/dL		
02	Clinical Pathology	Urine Full Report					
		Colour	Manual	-	-		
		Appearance					
		Specific Gravity	Strip method	Mission Strip	1.015-1.030		
		pH			5.0-9.0		
		Protein			Nil/Trace/ +/++/+++/ +++++		
		Glucose			Nil/Trace/ +/++/+++/ +++++		
		Ketone bodies			Nil/Trace/ +		
		Bilirubin			Nil/Trace/ +		
		Urobilinogen			Normal amounts /+		
		Nitrate			Nil/Trace/+		
		Pus cells			Microscopic	Olympus CX43	Counted Cells
		Red cells					Counted Cells
		Epithelial cells	+ / ++				
		Casts	Nil/ (casts type) + a few/ +				
		Crystals	Nil/ (crystal type) + a few/ +				

SI No	Field of Testing	Test	Test Method	Test Instrument	Linearity Range	
03	Haematology	Full Blood Count				
		RBC	DC detection method	Sysmex XS 1000i	0 - 8 X 10 ⁶ /μL	
		WBC	Flow cytometric method		Within ± 3% or ± 0.3 X 10 ³ / μL t 100 X 10 ³ /μL	
			Differential Count		Neutrophils - 0-99.9% Lymphocytes- 0 -99.9% Monocytes- 0-99.9% Eosinophils – 0-99.9% Basophils- 0-99.9%	
		HGB	SLS Hemoglobin method		0.0 – 25.0 g/dL	
		PLT	DC detection method		0.0 - 5000 X 10 ³ /μL	
		HCT	Calculated		0.0 - 60.0 HCT %	
		MCH				
		MCV				
		MCHC				
		RDW-CV				
		RBC	Electrical impedance method	Mindray BC 6200	(0 - 8.60) X10 ¹² /L	
		WBC	Light scattering method		(0 - 500) X 10 ⁹ /L	
		HGB	Colorimetric method		(0 - 260) g/L	
		PLT	Electrical impedance method		(0 - 5000) X 10 ⁹ /L	
		HCT	Calculated		(0 - 75) %	
		MCH	Calculated			
		MCV	Calculated			
		MCHC	Calculated			
		RDW-CV	RBC histogram			
		ESR	Westergern Method		2-120 mm	

Approved sample collection centres

SI no	Name of the sample collection center	Address	Matrix of the sample
01	Asiri Laboratory - Kandy	No 82, William Gopallawa Mawatha, Kandy	Plain, EDTA, Urine, Stool
02	Asiri Collection Centre - Bogambara	No 46/1, Keppetipola Road, Kandy	Plain, EDTA, Urine, Stool
03	Asiri Collection Centre - Gampola	No 36C, T.B Panabokke Mawatha, Gampola	Plain, EDTA, Urine, Stool
04	Asiri Collection Centre - Katugasthota	No 206A, Katugasthota Rd, Kandy	Plain, EDTA, Urine, Stool
05	Asiri Collection Centre – Kegalle	No 325, Main street, Kegalle	Plain, EDTA, Urine, Stool
06	Asiri Collection Centre - Mawanella	No 209, Colombo Road, Mawanella	Plain, EDTA, Urine, Stool
07	Asiri Collection Centre - Nawalapitiya	No 119, Kotmale Road, Nawalapitiya	Plain, EDTA, Urine, Stool
08	Asiri Collection Centre - Peradeniya	No 116, Peradeniya Road, Kandy	Plain, EDTA, Urine, Stool
09	Asiri Collection Centre - Pilimathalawa	No 197/1, Daulagala Road, Pilimathalawa	Plain, EDTA, Urine, Stool
10	Asiri Collection Centre - Thennekumbura	No 124, Thennekumbura, Kandy	Plain, EDTA, Urine, Stool
11	Asiri Collection Centre- Geliyoa	No 62, Kandy Rd, Geliyoa	Plain, EDTA, Urine, Stool
12	Asiri Collection Centre- Akurana	No 237/2/B, Matale Road, Akurana	Plain, EDTA, Urine, Stool
13	Asiri Collection Centre - Wattegama	No 45, Mathale Road, Waththegama	Plain, EDTA, Urine, Stool
14	Asiri Collection Centre - Dambulla	No 712, Anuradhapura Road, Dambulla	Plain, EDTA, Urine, Stool
15	Asiri Laboratories - Matale	No 91/A, Kandy Road, Matale	Plain, EDTA, Urine, Stool
16	Asiri Collection Center- Kundasale	No 58/2, Terence Garden, New Kandy Rd, Kundasale	Plain, EDTA, Urine, Stool
17	Asiri Collection Center- Madawala	No 50, Kandy Rd, Madawala Bazar, Madawala	Plain, EDTA, Urine, Stool

Acting Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment