

As an accredited laboratory, this laboratory is entitled to use the following accreditation symbol.



Valid from 22 June 2022
to 21 June 2025
Issued on 30 June 2022



ISO/ IEC 17025
TL 096-01

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 096-01

Food Quality Control Laboratory
RDHS Office Premises, S. W. R. D. Bandaranaike Mw,
Anuradhapura

**Scope of Accreditation: Performing Chemical Testing on Turmeric, Chilly
& Condiment powders**

The laboratory is accredited for the following tests;

Sl	Product(s) /Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
01	Turmeric Powder	Weight loss on oven dry at 105 °C for 3 hrs.	FQCL- AP-Oven-Re :00	8-22 %
		Total ash content	SLS 186 Part 3: 2008	6-12 %
		Acid insoluble ash content	SLS 186 Part 4: 2008	0.5-3.0 %

Sl	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed	Range of testing/ Limits of detection
02	Chilly Powder	Weight loss on oven dry at 105 °C for 3 hrs.	FQCL- AP-Oven-Re 00	10-22 %
		Total ash content	SLS 186 Part 3: 2008	6-12 %
		Acid insoluble ash content	SLS 186 Part 4: 2008	0.5-3.0 %
03	Condiment Powder	Weight loss on oven dry at 105 °C for 3 hrs.	FQCL- AP-Oven-Re :00	10-22 %
		Total ash content	SLS 186 Part 3: 2008	6-12 %
		Acid insoluble ash content	SLS 186 Part 4: 2008	0.5-3.0 %

Director /CEO
Sri Lanka Accreditation Board for Conformity Assessment