



Valid from to 30 June 2023
to 29 June 2026
Issued on 19 January 2024

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
TL 019-01

Schedule of Accreditation

Accreditation Scheme for Testing / Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: TL 019-01

Astron Ltd
No. 688, Galle Road
Ratmalana.

Scope of Accreditation: Performing Biological Testing on Pharmaceutical Liquid Preparations, Water, Milk Powder & Soya based products as per SLS and USP methods.

The Laboratory is accredited for the following tests appear on page 02;

Sl	Product(s) / Material of test	Specific tests performed	Test Method / Standard against which tests are performed (eg: xxx:2016)	Range of testing/ Limits of detection
Biological Testing				
1	Water	Coliform/100mL	SLS 614 : 2013 SLS 1461 - Part 1 Section 3	Nil to 1800 MPN
		<i>E.Coli</i> /100mL	SLS 614 : 2013 SLS 1461 - Part 1 Section 3	Nil to 1800 MPN
2	Milk Powder & Soya based products	Aerobic Plate Count/g or mL	SLS 731 : 2008 SLS 516 - Part 1 Section 1 USP - 44 NF - 41 - 2023 Section 2021	10 to 5×10^4 cfu/g or mL 10 to 3×10^3 cfu/g or mL
		Coliform/g or mL	SLS 731 : 2008 SLS 516 – Part 3 Section 1	Nil to 110 MPN
		<i>Staphylococcus aureus</i> / g or mL	USP - 38 NF - 33 : 2015 Section 2022	Present or Absent
		<i>E. Coli</i> / g or mL	SLS 731 : 2008 SLS 516 - Part 12 : 2013 USP - 46 NF - 41 : 2023 Section 2022	Nil to 110 MPN Present or Absent
3	Pharmaceutical Liquid Preparations	Aerobic Plate Count / g or mL	USP - 46 NF - 41 - 2023: Section 2021	10 to 3×10^3 cfu/g or mL
		<i>E. Coli</i> / g or mL	USP - 46 NF - 41 - 2023: Section 2022	Present or Absent
		<i>Staphylococcus aureus</i> / g or mL	USP -38 NF-33-2015: Section 2022	Present or Absent

Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment