



Valid from 11 September 2024
To 27 December 2026
Issued on 11 September 2024

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/ IEC 17025
CL 014-02

Schedule of Accreditation

Accreditation Scheme for Testing Laboratories
Sri Lanka Accreditation Board for Conformity Assessment
Accreditation Number: CL 014-02

**Dimension Laboratory,
Measurement Units Standards & Services Department
Mahenawatte, Pitipana
Homagama**

Scope of Accreditation: : Performing Length Calibration as per calibration method appearing
in this schedule.

The laboratory is accredited for the following calibration as per given in the page 01 of 01.

SI No	Type of Instrument / Gauge	Calibration performed/ Measured Quantity	Calibration methods / Measurement procedure	Range of calibration	Readability/ Resolution as applicable	Calibration Measurement Capability (Expanded uncertainty at a coverage probability of 95 %; coverage factor of k = 2.)	Location (Site/ In house)
1	External Micrometer	Comparison with gauge blocks	CP2: DM/CP/EM/D2	0 – 25 mm	0.001 mm	$\sqrt{(0.83 \mu\text{m})^2 + (1.7 \times 10^{-6} \text{ } l)^2}$	In house

Director/CEO (Coverup Duties)
Sri Lanka Accreditation Board for Conformity Assessment