



Valid from 28 May 2025
to 27 May 2029
Issued on 04 July 2025

As an accredited laboratory, this laboratory is entitled to
use the following accreditation symbol.



ISO/IEC 17025
CL 028- 01

Schedule of Accreditation

Accreditation Scheme for Testing/Calibration Laboratories
Sri Lanka Accreditation Board for Conformity Assessment

Accreditation Number: CL 028-01

Metro Calibration Services (Pvt.) Ltd

No. 83/1, Galle Road,
Mount Lavinia.

Scope of Accreditation: Performing Calibrations in Mass and related Quantities (Pressure calibrations) as per the calibration methods appearing in this Schedule.

The laboratory is accredited for the following calibrations as per given from page 02 to 02.

| SI No | Type of Instrument / Gauge | Calibration performed/ Measured Quantity | Calibration methods / Measurement procedure | Range of calibration | Readability/ Resolution as applicable | Calibration Measurement Capability (Expanded uncertainty at a coverage probability of 95 %; coverage factor of k = 2.) | Location (Site/ In house) |
|--------------|--|---|---|-----------------------------|--|---|----------------------------------|
| 01 | Bourdon Tube Hydraulic Pressure Gauges | Guage Calibration | Calibration of Pressure guages DKD-R6-1:2014 Guideline and MCS-PRO-12 | 0 – 400 bar | 0.5 bar | 1.5 bar | On site / In House |

Acting Director/CEO
Sri Lanka Accreditation Board for Conformity Assessment